

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
01C	Controllo	309106	NI 8000	Nimex	Multimetro digitale / Digital multimeter	2022/06/10	Annuale - Yearly	10/06/2021	
01R	Riferimento	1370/2/D	F2005	Paint Test Equipment	Pellicole di calibrazione / Calibration foils		n.a.	5048	
02C	Controllo	C299	Eban 3000 MK2	Paint Test Equipment	Spessimetro / Coating thickness meter		Prima di ogni utilizzo - Before each use	107225	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
02R	Riferimento	4500-6-I	--	Paint Test Equipment	Fogli di calibrazione / Calibration Foils		n.a.	107224	
03C	Controllo	3/2/V	Snap Gauge	Paint Test Equipment	Replica testex tape	2022/09/17	Annuale - Yearly	17/09/2021	
03R	Riferimento	-	ISO 8501-1	ISO	Standard visivo con foto / Visual standard with photos		n.a.		
04C	Controllo	353/2/V	S3002	Paint Test Equipment	Strumento per prove di continuità a spugna umida / Pintech wet sponge pinhole tester	2022/03/25	Annuale - Yearly	7728	
04R	Riferimento	04533	W5001	Paint Test Equipment	Spessimetri ad umido usa e getta 50-900 microns / Plastic disposable wet film thickness gauge 50-900 microns		n.a.	112193	disposable instrument
05C	Controllo	S/N AT02374	Positest	Defelsko	Strumento per prove di adesione digitale / Digital Pull-Off Adhesion Tester	2023/01/10	Annuale - Yearly	10/01/2022	
05R	Riferimento	04394	R1002	Paint Test Equipment	Testex Tape X Coarse / Testex Tape X Coarse		n.a.	111804	Testex Tape X Coarse for Replica Tape Test 03C
06C	Controllo	-	X2002	Paint Test Equipment	Quadrettatore per prove di aderenza / Cross hatch cutter		Giornaliera - Daily	4045/02/V	verification before each use
07C	Controllo	--	P2004	Paint Test Equipment	Conductivity Meter of Bresle Salt Test Kit		Giornaliera - Daily	4043/02/V	verification before each use
07R	Riferimento	04534	PS005	Paint Test Equipment	Soluzione per calibrazione conduttimetro / Conductivity Meter calibration Solution		n.a.	112194	

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
08C	Controllo	375/5/B	S4003	Paint Test Equipment	Holiday Tester		Giornaliera - Daily	114780	Verification before each use
08I	riferimento	--		CMV	Termocoppie forno CMV - Thermocouples oven CMV	2022/07/10	Semestrale - six months	10/01/2022	
09C	Controllo	VQ350008	Shore D	PCE Group	Durometro per prove Shore D / Durometer for Shore D tests		Giornaliera - Daily	D-72644	verification before each use
09R	Riferimento - Reference	181211131	SA6510TD	Sama Tools	Campioni certificati Shore D - Shore D Calibration Samples		n.a.	181211131	
10C	Controllo	06013041	HD 2101.2	Delta Ohm	Igrotermometro digitale / Digital hygrothermometer	2022/07/17	Annuale - Yearly	17/07/2021	
11C	Controllo	--	XA001	--	Coltello e nastro adesivo per prove di adesione / Knife and tape for adhesion test		Giornaliera - Daily	4044/02/V	verification before each use
12C	Controllo	Y132260	DIN / BIL / M30	STROPS	Bilancia per pitture - Balance for paints	2023/01/29	Annuale - Yearly	29/01/2022	
13C	Controllo	12/9/W	MT	Paint Test Equipment	Termometro magnetico / Magnetic thermometer	2022/09/07	Annuale - Yearly	07/09/2021	
14 R	Riferimento	5884/7/L	F2005	Paint Test Equipment	Fogli di calibrazione / Calibration foils		n.a.	114779	
14C	Controllo	202/3/Z	Eban 4000	Paint Test Equipment	Spessimetro / Coating thickness meter		Prima di ogni utilizzo - Before each use	104445	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
15C	Controllo	SN090009	2049	Amel Instruments	Apparecchiatura per distacco catodico /Cathodic disbondment equipment	2022/07/17	Annuale - Yearly	17/07/2021	
15R	Riferimento	1711/9/D	F2005	Paint Test Equipment	Spessori certificati / Calibration Foils		n.a.	6472	

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
16C	Controllo	70000984	Load cell 60001 2.5T - 3000 M	VIP - PAVONE SISTEMI OIML	Apparecchiatura per prova di pelatura / Peeling test equipment	2023/01/29	Annuale - Yearly	29/01/2022	
16R	Riferimento	--	FV001	Paint Test Equipment	Spessori certificati / Calibration Foils		n.a.	5805/3/Z	
17+36	Controllo	--	--	--	Pirometro ottico+ indicatori di temperatura Tempilstick / Optical pyrometer+Tempilstick temperature indicators		--	--	
17C	Controllo	07050789	PCE 889	PCE Group	Pirometro ottico - Optical pyrometer	2022/07/09	annuale- yearly	09-07-2021	
17R	Riferimento	D7070	--	Pasquali Balance	Massa da 20 kg. / Sample weight 20 kg.		n.a.	2060/09	
18C	Controllo	--	Dn 25 x 7000	VIP	Strumento per prove di impatto su risers / Impact tester for risers		Giornaliera - Daily		verification before each use
18R	Riferimento			Mettler Toledo	Peso a campione / Sample weight		n.a.		
19C	Controllo	--	MAC 0-20	Magnetic Analysis Corporation	Gaussmetro / Gaussmeter		Giornaliera - Daily		verification before each use
20C	Controllo	--	Dn 25 / Dn 15.9	VIP	Strumento per prove di impatto / Impact tester		Giornaliera - Daily		verification before each use
21C	Controllo	A50050	S.900.1000	M.CONTI	Calibro meccanico da 0 a 1000 mm / Slide gauge from 0 to 1000 mm	2031/02/12	Dieci anni - Ten years	V1102003D+	
21I	riferimento	--		Soprim	Termocoppia forno Soprim - Thermocouple oven Soprim	2022/07/10	Semestrale - Six months	10/01/2022	
22C	Controllo	103010077601 7	AR 30	VIP - STAB	Apparecchiatura per distacco catodico /Cathodic disbondment equipment	2022/07/17	Annuale - Yearly	17/07/2021	
23C	Controllo	--	--	--	Nastro trasparente - Transparent Tape (Dust comparator chart)		Giornaliera - Daily	108450	verification before each use

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
24C	Controllo	--	--	Riveco Generalsider	Strumento per misurazione durezza Buchholz / Gauge for hardness test Buchholz		n.a.		Instrument on loan
26+36	Controllo				Temometro digitale a contatto + Tempilstik / Contact digital thermometer + Tempilstik				
26C	Controllo	1710/5/G	T2003	Paint Test Equipment	Termometro digitale / Digital Thermometer	2022/05/04	Annuale - Yearly	04/05/2021	
27C	Controllo	123		Elcometer	Rugosimetro / Surface roughness tester				
28C	Controllo	034-14184A	Deltascope	Fischer	Spessimetro / Coating thickness meter		Prima di ogni utilizzo - Before each use		Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
29C	Controllo	533076	6210	SRT	Rugosimetro / Surface Roughness Tester	2023/01/29	Annuale - Yearly	29/01/2022	
29R	Riferimento			n.a.	campione per calibrazione rugosimetro / sample for calibration surface roughness tester		n.a.		
30C	Controllo	K2011		VIP - FERVI	Comparatore centesimale / Centesimal comparator	2022/03/12	Annuale - Yearly	12/03/2021	verification before each use
31C	Controllo	2P6632113262	CM8801FN	Nicetu	Spessimetro ad ultrasuoni / Ultrasonic Coating thickness Gauge		Prima di ogni utilizzo - Before each use		Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
32C	Controllo	110002587	PHASCOP E PMP10	Helmut Fischer	Spessimetro speciale per rivestimenti di TSA su SS / Special coating thickness meter for TSA on SS		Prima di ogni utilizzo - Before each use	3501	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
33C	Controllo	1015085	XJP-H100	AmScope	Trinocular Metallurgical Microscope + 1.3MP USB Camera	2022/09/17	Annuale - Yearly	17-09-2021	verification before each use
34C	Controllo	MY47004057 - B214826218	FP90 + FP85HT	Mettler Toledo	Calorimetro a scansione differenziale / Differential Scanning Calorimeter	2023/01/10	Annuale - Yearly	10/01/2022	

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
34R	Riferimento	BV335286	0319-0033	Perkin Elmer	Pillole di Indio / Indium pills		n.a.	11-03-2021	
35C	Riferimento	--	W2001	Paint Test Equipment	Spessimetri ad umido 25-1500 microns / Wet film thickness gauge 25-1500 microns		Giornaliera - Daily	1126	verification before each use
36C	Controllo	-	-	-	Indicatori di temperatura Tempilstik / Tempilstik temperature indicators				
37C	Controllo	GM07-1296	GM07	HIRST	Gaussmetro / Gaussmeter	2022/05/26	Cinque anni - Five years	5214/17	
38C	Controllo				Macchina per curvatura manuale / Manual bending machine				
40/C	Controllo	159/5/C	E2001	Paint Test Equipment	Jeep Meter	2022/06/06	Cinque anni - Five years	101581	
41/C	Controllo	W1205656	KB 120-3N	Kern	Bilancia di precisione / Precision Balance				
42/C	Controllo	1272/2/D	C4005	Paint Test Equipment	Spessimetro per acciaio inossidabile / Coating thickness meter for stainless steel		Prima di ogni utilizzo - Before each use	5044	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
42R	Riferimento	1368/2/D	F2001	Paint Test Equipment	Spessori certificati / Calibration Foils		n.a.	5043	
43C	Controllo	70000984	60001 2.5T - 3000 M	Pavone Sistem - OIML	Cella di carico portata massima 2500 kg / Load cell maximum load 2500 kg.	2023/01/29	Annuale - Yearly	29/01/2022	
44C	Controllo	8750v3733799 1	HS301	Paint Test Equipment	Strumento per calcolo Dew point / Dew point meter	2022/07/17	Annuale - Yearly	17/07/2021	
45/C	Controllo	05351109	Art. P001/04	FERVI - VEPRUG Srl	Pressa idraulica per prove di compressione e piegatura 51 MPa / Hydraulic Press for compression test and bending 51 MPa		--	--	
46 C	controllo	761		Paint Test Equipment	Potassium Ferricyanide Test Papers		n.a.	4186	

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
 NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
47C	Controllo		R2006	Paint Test Equipment	Comparatore di rugosità / Roughness comparator ISO 8503-1/2		n.a.	2085	
48C	Controllo	--	741150	Stalgast	Bagno Termostatico / Thermostatic Bath		--	--	
49C	Controllo	001466	LE438-MA 130	Mettler Toledo	Elettrodo a combinazione pH / pH Combination electrode				
50C	Controllo		741150	Stalgast	Bagno termostato / Thermostatically Bath				
51C	Controllo	Y121027	DIN / BIL / M30	STROPS	Bilancia elettronica / Electronic Balance	2023/01/29	Annuale - Yearly	29/01/2022	
52C	Produzione	022/08	FS	KERN & SON	Cella di carico portata massima 50 kg / Load cell maximum load 50 kg.	2023/01/10	Annuale - Yearly	10/01/2022	
53C	Controllo	--	1XC28500	G3 Ferrari Group	Forno statico / Soaking oven		n.a.	n.a.	
54C	Controllo	10129547	HTI 3301	HT Italia	Pirometro ottico / Optical Pyrometer	2022/07/17	Annuale - Yearly	17/07/2021	
55C	Controllo		TM-8812	STC S.r.l.	Misuratore di spessori ad ultrasuoni / Ultrasonic thickness gauge				
56C	Controllo			VIP	Strumento per prove di impatto 400-500-600-700J / Impact Tester 400-500-600-700J		Giornaliera - Daily		verification before each use
57C	Controllo	n.a.	P2004	Paint Test Equipment	Conductivity Meter		Giornaliera - Daily	4043/02/V	verification before each use
58C	Controllo				Contact digital thermometer + Tempilstik + digital pyrometer				Vd. / See 26C + 36C + 17C
59C	Controllo	I.336757	TM-947 SD	Lutron	Termometro a 4 canali - 4 channels thermometer - Datalogger with thermocouples	2022/09/07	Annuale - Yearly	07/09/2021	
60C	Controllo	WLI034		BAKER	Comparatore centesimale / Centesimal comparator	2022/03/12	Annuale - Yearly	12-03-2021	verification before each use

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
61C	Controllo	31090478WS	115	FLUKE	Multimetro - Multimeter	2025/07/22	Dieci anni - Ten years	LAT 019 43715	
62C	Controllo	C439	CS303	Paint Test Equipment	Spessimetro per rivestimento - Coating thickness meter 1-20.0mm		Prima di ogni utilizzo - Before each use	108702	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
62R	Controllo / Control	5035/7/J	FV001	Paint Test Equipment	Spessori certificati / Calibration foils		--	108703	
63C	Controllo	B30FZKB0N00 QLF3Q0137	BD-103RA A	Haier	Congelatore - Chest Freezer		--	--	per prove a bassa temperatura - for tests at low temperature
64C	Controllo	70000984	Load cell 60001 2.5T - 3000 M	VIP - PAVONE SISTEMI OIML	Apparecchiatura per prova di allungamento / Elongation test equipment	2023/01/29	Annuale - Yearly	29/01/2022	
66c	Controllo	--	0-150 mm/ 0-6 inch	VERNIER	Calibro digitale - Digital Calliper (0-150 mm/ 0-6 inch)	2022/09/07	Annuale	07/09/2021	
67C	Controllo	SN140019	Mod. 2549	Amel Instruments	Apparecchiatura per distacco catodico / Cathodic disbondment equipment	2022/10/19	Annuale - Yearly	19/10/2021	
68C	Controllo - Control		CS215010 46	TQC	Durometro Bucholz - Bucholz Durometer	2025/12/14	5 anni - 5 years	SP1900-408	
69C	Controllo	34	VT60	Wunder	Bilancia per pitture - Balance for paint	2022/03/12	Annuale - Yearly	12-03-2021	
70C	Controllo	--	Several	Several	Metro / Tape measure		Giornaliera / Daily	--	verification the integrity before each use
71C	Controllo	--	P2004	Paint Test Equipment	Visual and Conductivity Meter of Bresle Salt Test Kit		Giornaliera - Daily	4043/02/V	verification before each use
72C	Controllo	130001944	Dualscope MP0RH-FP	Helmut Fischer	Spessimetro - Coating thickness meter		Prima di ogni utilizzo - Before each use	1421-2019	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
72R	--	00100	W3001	Paint Test Equipment	Spessimetri ad umido / Wet film thickness gauge		Giornaliera / Daily	100265	
73C	Controllo	--		--	Coltello - Knife (Stanley type of similar)		Giornaliera - Daily	--	verification before each use
73R	Riferimento	5883/7/L	F2001	Paint Test Equipment	Spessori certificati / Calibration Foils		--	114778	
74C	Controllo	02091704	T050	FERVI	Multimetro Digitale / Digital Multimeter	2022/03/12	1 anno / 1 year	12-03-2021	
74R		--	W4002	Paint Test Equipment	Spessimetri ad umido / Wet film thickness gauge		Giornaliera / Daily	106107	
75R	Riferimento	603-479 / 603-872	KAL-N 75micron / KAL-N 500micron	Helmut Fischer	Fogli di calibrazione - Calibration Foils		--	141-2017	
76C	Controllo - Control	1013-5-I	S4003	Paint Test Equipment	Holiday Tester - Holiday Tester		Giornaliera - Daily	106121	
78C	Controllo	1701738		TC Direct	Termometro Digitale - Digital Thermometer	2022/07/26	1 anno - 1 year	26-07-21	
79C	Controllo	C1073		Paint Test Equipment	Spessimetro - Coating thickness meter		Prima di ogni utilizzo - Before each use	106613	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report
79R	Riferimento	4703/11/I		Paint Test Equipment	Spessori Certificati / Calibration Foils		--	106612	
80C	Giornaliero - Daily	010058401	EL-USB-T C-LCD	LASCAR Electronics	Thermocouple Temperature Data Logger	2023/01/29	Annuale - Yearly	29/01/2022	
81C	Giornaliero - Daily	010045223	EL-USB-2-LCD+	LASCAR Electronics	High Accuracy Temp & RH Data Logger	2023/01/29	Annuale - Yearly	29/01/2022	
82C	Giornaliero / Daily	010058788	EL-USB-T C-LCD	LASCAR Electronics	Registratore di dati di temperatura a termocoppia / Thermocouple temperature data Logger	2023/02/08	Annuale / Yearly	08-02-2022	

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
83C	Giornaliero / Daily	010045466	EL-USB-2-LCD+	LASCAR Electronics	Registratore di dati Temperatura e RH / High accuracy temperature & RH data Logger	2023/02/08	Annuale / Yearly	08-02-2022	
84C	Controllo	560-10D	Shore D	Gainexpress	Durometro per prove Shore D - Durometer for Shore D tests		Giornaliera - Daily	181211131	
85C	Controllo / Control	S565570	T064	Fervi	Pirometro ottico / Optical Pyrometer	2022/08/30	Annuale / Yearly	08001806	
87C	Controllo	DE2875637940747	PH-01	Dustgo	Misuratore di Ph / Portable Ph meter	2022/04/14	Annuale / Yearly	14/04/2021	
88C	Controllo	DE2875637940747	EC-02	Dustgo	Conduttimetro / Portable TDS Meter	2022/04/14	Annuale / Yearly	14/04/2021	
89C	Controllo / Control	560-10D	LX-D	Gain Express	Durometro / Durometer	2031/04/27	10 anni / 10 years	28-04-2021	
89R	Riferimento / Reference	200714785	Type D	Yfyiqi	Campioni certificati Shore D / Shore D Calibration Samples		n.a.	n.a.	
90C	Controllo / Control	560-10A	LX-A	Gain Express	Durometro / Durometer	2031/04/27	10 anni / 10 years	28-04-2021	
90R	Riferimento / Reference	200915833	Type A	Yfyiqi	Campioni certificati Shore A / Shore A Calibration Samples		n.a.	n.a.	
91C	Controllo / Control	n.a.	SP3000	TQC Sheen	Prove di Adesione e Preparazione acciaio / Master Paint Plate Multi-Teste		n.a.	n.a.	
91R	Riferimento / Reference	5378	HI7031L	Hanna Instruments	Soluzione calibrazione Conduttimetro / Conductivity Meter calibration solution	2025/05/12	5 anni / 5 years	12E04	
92C	Riferimento		KZ2000-10M	KZYEE	Endoscopio 10 metri - Endoscope 10 meters		Giornaliera / Daily	N/A	
93C	Controllo	C2295	Eban C5002	Paint Test Equipment	Spessimetro / Coating thickness meter		Prima di ogni utilizzo - Before each use	115841	Calibration, verification of accuracy and adjustment with certified sample to be indicated on report

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status

Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
93R	Riferimento	6036/1/M	F2005	Paint Test Equipment	Fogli di calibrazione / Calibration foils		n.a.	115840	
94C	Controllo	POC-1519	Mega Check Pocket FN	ASSI Control	Spessimetro / Coating thickness meter		Prima di ogni utilizzo - Before each use	619/22	
el					Laboratorio esterno / External laboratory				
Ham					Martello / Hammer				
na					Non applicabile / Not applicable				
PC01	Controllo / Control	C2053 - C2054	C5006	Paint Test Equipment	Spessimetro / Coating thickness meter		Prossimo utilizzo / Next use	108441 - 108442	
PC02	Controllo / Control	6224926-5808 A04038	H4001	Paint Test Equipment	Calcolatore Dew Point / Dew Point Meter	2021/06/18	1 anno / 1 year	108444 - 108445	
PC03	Controllo / Control	2050/6/J	R1004	Paint Test Equipment	Rugosimetro / Surface Roughness Tester	2021/06/18	1 anno / 1 year	108443	
PC04	Controllo / Control	03020 - 828/6/J	P2005	Paint Test Equipment	Conduttivimetro Bresle Salt Test Kit / Conductivity Meter Bresle Salt Test Kit		Prossimo utilizzo / Next use	108446 - 108448	
PC05	Controllo / Control	03024	W4001 - W4002	Paint Test Equipment	Spessimetri ad umido / Wet film thickness gauge		Prossimo utilizzo / Next use	108452	
PC06	Controllo / Control	03022 - 03023	PS201 - PS202	Paint Test Equipment	Nastro e Grafico per prova polvere / Dust Test Tape and Chart		Prossimo utilizzo / Next use	108450 - 108451	
PR01	Calibrazione e / Calibration	4995/6/J	F2001	Paint Test Equipment	Fogli di calibrazione / Calibration foils		n.a.	108440	
PR03	Calibrazione e / Calibration	03019	R1002	Paint Test Equipment	Fogli di spessori certificati / Certified thicknesses foils		n.a.	108447	

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
 NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued

Measuring Instruments Status



Pos.	Use	Serial Number	Type	Manufacturer	Description	Expiry	Frequency	Certificate	Notes
PR04	Calibrazione e / Calibration	03021	PS005	Paint Test Equipment	Soluzione di calibrazione conduttimetro / Conductivity Meter Calibration Solution		n.a.	108449	
tbc					Strumentazione da calibrare / Instrumentation to be calibrated				
VisDo	Controllo				Visivo documenti / Visual documents				
VisIO	Controllo				Visivo accordo istruzione operativa / Visual according operative instruction				
VisST	Controllo				Visivo accordo standard / Visual according standard				

Approvato dal Responsabile Assicurazione Qualità / Approved by Quality Assurance Manager

Pietro Viganò
NACE Coating Inspector Level 2 - Certified

Il presente certificato non viene firmato perché generato elettronicamente / The present certificate is not signed because electronically issued